SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 12/09/2019

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:51
INTERNAL STANDARD	VERIFIED	12:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	12:51
BLANK TEST	0.00	12:52
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	12:53
BLANK TEST	0.00	12:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	12:54
BLANK TEST	0.00	12:55







SOLUTION CHANGE

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 12/02/2019

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:30
INTERNAL STANDARD	VERIFIED	10:30
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:31
BLANK TEST	0.00	10:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:32
BLANK TEST	0.00	10:33
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:33
BLANK TEST	0.00	10:34







SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 11/25/2019

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:48
INTERNAL STANDARD	VERIFIED	20:48
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:48
BLANK TEST	0.00	20:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:50
BLANK TEST	0.00	20:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:51
BLANK TEST	0.00	20:52



SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 11/18/2019

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:37
INTERNAL STANDARD	VERIFIED	17:37
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:38
BLANK TEST	0.00	17:39
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:39
BLANK TEST	0.00	17:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:40
BLANK TEST	0.00	17:41







SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 11/11/2019

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:58
INTERNAL STANDARD	VERIFIED	07:59
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:59
BLANK TEST	0.00	08:00
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:00
BLANK TEST	0.00	08:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:01
BLANK TEST	0.00	08:02







SOLUTION CHANGE

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 11/04/2019

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:20
INTERNAL STANDARD	VERIFIED	23:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:21
BLANK TEST	0.00	23:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:22
BLANK TEST	0.00	23:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:23
BLANK TEST	0.00	23:24







SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 11/04/2019

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:29
INTERNAL STANDARD	VERIFIED	20:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:30
BLANK TEST	0.00	20:31
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:31
BLANK TEST	0.00	20:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:32
BLANK TEST	0.00	20:33







SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 10/28/2019

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:03
INTERNAL STANDARD	VERIFIED	18:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:04
BLANK TEST	0.00	18:05
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:05
BLANK TEST	0.00	18:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:06
BLANK TEST	0.00	18:07







SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 10/21/2019

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:53
INTERNAL STANDARD	VERIFIED	19:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:54
BLANK TEST	0.00	19:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:55
BLANK TEST	0.00	19:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:56
BLANK TEST	0.00	19:57







SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 10/14/2019

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:13
INTERNAL STANDARD	VERIFIED	20:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:14
BLANK TEST	0.00	20:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:15
BLANK TEST	0.00	20:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:16
BLANK TEST	0.00	20:17







SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 10/07/2019

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:54
INTERNAL STANDARD	VERIFIED	20:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:54
BLANK TEST	0.00	20:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:55
BLANK TEST	0.00	20:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:57
BLANK TEST	0.00	20:58







SOLUTION CHANGE

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 10/01/2019

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:24
INTERNAL STANDARD	VERIFIED	20:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:25
BLANK TEST	0.00	20:26
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:26
BLANK TEST	0.00	20:27
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:27
BLANK TEST	0.00	20:28







SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 09/23/2019

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:10
INTERNAL STANDARD	VERIFIED	20:10
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:11
BLANK TEST	0.00	20:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:12
BLANK TEST	0.00	20:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:13
BLANK TEST	0.00	20:14







SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 09/16/2019

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:56
INTERNAL STANDARD	VERIFIED	21:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:57
BLANK TEST	0.00	21:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:58
BLANK TEST	0.00	21:59
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:59
BLANK TEST	0.00	22:00







SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 09/09/2019

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:30
INTERNAL STANDARD	VERIFIED	20:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:31
BLANK TEST	0.00	20:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:32
BLANK TEST	0.00	20:33
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:33
BLANK TEST	0.00	20:34







SOLUTION CHANGE

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 09/04/2019

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:21
INTERNAL STANDARD	VERIFIED	01:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:21
BLANK TEST	0.00	01:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:22
BLANK TEST	0.00	01:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:24
BLANK TEST	0.00	01:25







SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 08/26/2019

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:59
INTERNAL STANDARD	VERIFIED	19:59
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:00
BLANK TEST	0.00	20:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:01
BLANK TEST	0.00	20:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:02
BLANK TEST	0.00	20:03







SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 08/19/2019

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:40
INTERNAL STANDARD	VERIFIED	09:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:41
BLANK TEST	0.00	09:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:42
BLANK TEST	0.00	09:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:43
BLANK TEST	0.00	09:44







SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 08/12/2019

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:33
INTERNAL STANDARD	VERIFIED	19:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:34
BLANK TEST	0.00	19:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:35
BLANK TEST	0.00	19:36
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:36
BLANK TEST	0.00	19:37





SOLUTION CHANGE

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 08/05/2019

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:41
INTERNAL STANDARD	VERIFIED	18:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:41
BLANK TEST	0.00	18:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:43
BLANK TEST	0.00	18:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:44
BLANK TEST	0.00	18:45







SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 07/29/2019

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:34
INTERNAL STANDARD	VERIFIED	18:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:34
BLANK TEST	0.00	18:35
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:35
BLANK TEST	0.00	18:36
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:37
BLANK TEST	0.00	18:38



SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 07/22/2019

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:27
INTERNAL STANDARD	VERIFIED	18:27
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:28
BLANK TEST	0.00	18:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:29
BLANK TEST	0.00	18:30
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:30
BLANK TEST	0.00	18:31







SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 07/15/2019

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:19
INTERNAL STANDARD	VERIFIED	21:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:20
BLANK TEST	0.00	21:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:21
BLANK TEST	0.00	21:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:22
BLANK TEST	0.00	21:23







SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 07/08/2019

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:52
INTERNAL STANDARD	VERIFIED	21:52
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:53
BLANK TEST	0.00	21:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:54
BLANK TEST	0.00	21:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:55
BLANK TEST	0.00	21:56







SOLUTION CHANGE

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 07/01/2019

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:40
INTERNAL STANDARD	VERIFIED	19:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:41
BLANK TEST	0.00	19:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:42
BLANK TEST	0.00	19:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:43
BLANK TEST	0.00	19:44





SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 07/01/2019

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:46
INTERNAL STANDARD	VERIFIED	18:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:46
BLANK TEST	0.00	18:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:47
BLANK TEST	0.00	18:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:49
BLANK TEST	0.00	18:50