Granville
PD

2019

Weekly Test

SOLUTION CHANGE

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

)ate: 12/07/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:37
INTERNAL STANDARD	VERIFIED	22:37
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:38
BLANK TEST	0.00	22:38
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:39
BLANK TEST	0.00	22:39
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:40
BLANK TEST	0.00	22:41







SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 11/30/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:32
INTERNAL STANDARD	VERIFIED	18:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:32
BLANK TEST	0.00	18:33
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:33
BLANK TEST	0.00	18:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:34
BLANK TEST	0.00	18:35







SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 11/24/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:43
INTERNAL STANDARD	VERIFIED	21:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	21:43
BLANK TEST	0.00	21:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:44
BLANK TEST	0.00	21:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:45
BLANK TEST	0.00	21:46







SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 11/18/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:24
INTERNAL STANDARD	VERIFIED	01:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:25
BLANK TEST	0.00	01:26
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:26
BLANK TEST	0.00	01:27
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:27
BLANK TEST	0.00	01:28







SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 11/10/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:51
INTERNAL STANDARD	VERIFIED	02:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:52
BLANK TEST	0.00	02:52
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:53
BLANK TEST	0.00	02:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:54
BLANK TEST	0.00	02:55







SOLUTION CHANGE

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 11/03/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:09
INTERNAL STANDARD	VERIFIED	04:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:09
BLANK TEST	0.00	04:10
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:11
BLANK TEST	0.00	04:11
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:12
BLANK TEST	0.00	04:13



SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 10/27/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:50
INTERNAL STANDARD	VERIFIED	05:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:50
BLANK TEST	0.00	05:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:52
BLANK TEST	0.00	05:52
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:53
BLANK TEST	0.00	05:53







SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 10/20/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:39
INTERNAL STANDARD	VERIFIED	03:39
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:39
BLANK TEST	0.00	03:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:41
BLANK TEST	0.00	03:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:42
BLANK TEST	0.00	03:43





SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE LEPT
DATAMASTER dmt: 113306

Date: 10/13/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:11
INTERNAL STANDARD	VERIFIED	00:11
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:12
BLANK TEST	0.00	00:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:13
BLANK TEST	0.00	00:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:14
BLANK TEST	0.00	00:15







SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 09/29/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:05
INTERNAL STANDARD	VERIFIED	01:05
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:05
BLANK TEST	0.00	01:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:06
BLANK TEST	0.00	01:07
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:08
BLANK TEST	0.00	01:08





SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 09/24/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:38
INTERNAL STANDARD	VERIFIED	10:38
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:39
BLANK TEST	0.00	10:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:40
BLANK TEST	0.00	10:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:41
BLANK TEST	0.00	10:42



SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 09/14/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:23
INTERNAL STANDARD	VERIFIED	21:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:23
BLANK TEST	0.00	21:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:25
BLANK TEST	0.00	21:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:26
BLANK TEST	0.00	21:26





SOLUTION CHANGE

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 09/07/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:20
INTERNAL STANDARD	VERIFIED	20:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:20
BLANK TEST	0.00	20:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:21
BLANK TEST	0.00	20:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:22
BLANK TEST	0.00	20:23





SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 09/01/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:57
INTERNAL STANDARD	VERIFIED	19:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:58
BLANK TEST	0.00	19:59
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:59
BLANK TEST	0.00	20:00
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:00
BLANK TEST	0.00	20:01







SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 08/24/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:25
INTERNAL STANDARD	VERIFIED	05:26
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:26
BLANK TEST	0.00	05:27
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:27
BLANK TEST	0.00	05:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:28
BLANK TEST	0.00	05:29







SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 08/17/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:55
INTERNAL STANDARD	VERIFIED	04:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:56
BLANK TEST	0.00	04:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:57
BLANK TEST	0.00	04:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:58
BLANK TEST	0.00	04:59





SOLUTION CHANGE

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 08/10/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:20
INTERNAL STANDARD	VERIFIED	05:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:20
BLANK TEST	0.00	05:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:21
BLANK TEST	0.00	05:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:22
BLANK TEST	0.00	05:23



SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 08/03/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:20
INTERNAL STANDARD	VERIFIED	06:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:21
BLANK TEST	0.00	06:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:22
BLANK TEST	0.00	06:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:23
BLANK TEST	0.00	06:24



3W?v☺afn@☺!

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 07/27/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:07
INTERNAL STANDARD	VERIFIED	04:07
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:08
BLANK TEST	0.00	04:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:09
BLANK TEST	0.00	04:10
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:10
BLANK TEST	0.00	04:11







SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 07/22/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:02
INTERNAL STANDARD	VERIFIED	22:02
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:03
BLANK TEST	0.00	22:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:04
BLANK TEST	0.00	22:05
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:05
BLANK TEST	0.00	22:06





SOLUTION CHANGE

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 07/19/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:45
INTERNAL STANDARD	VERIFIED	14:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:45
BLANK TEST	0.00	14:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:47
BLANK TEST	0.00	14:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:48
BLANK TEST	0.00	14:48





SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 07/13/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:40
INTERNAL STANDARD	VERIFIED	19:40
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	19:40
BLANK TEST	0.00	19:41
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	19:41
BLANK TEST	0.00	19:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	19:43
BLANK TEST	0.00	19:43



SOLUTION CHANGE

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 07/07/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:29
INTERNAL STANDARD	VERIFIED	03:29
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:30
BLANK TEST	0.00	03:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	03:31
BLANK TEST	0.00	03:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	03:32
BLANK TEST	0.00	03:33







SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 06/30/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:56
INTERNAL STANDARD	VERIFIED	02:56
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:57
BLANK TEST	0.00	02:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:58
BLANK TEST	0.00	02:59
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:59
BLANK TEST	0.00	03:00





SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 06/22/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:04
INTERNAL STANDARD	VERIFIED	22:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:04
BLANK TEST	0.00	22:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:06
BLANK TEST	0.00	22:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:07
BLANK TEST	0.00	22:08





SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 06/15/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:57
INTERNAL STANDARD	VERIFIED	05:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:57
BLANK TEST	0.00	05:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:59
BLANK TEST	0.00	05:59
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:00
BLANK TEST	0.00	06:01





SOLUTION CHANGE

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 06/12/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:15
INTERNAL STANDARD	VERIFIED	21:15
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:15
BLANK TEST	0.00	21:16
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:16
BLANK TEST	0.00	21:17
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:18
BLANK TEST	0.00	21:18





SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 05/26/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 19030

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:48
INTERNAL STANDARD	VERIFIED	04:48
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:49
BLANK TEST	0.00	04:50
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:50
BLANK TEST	0.00	04:51
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:51
BLANK TEST	0.00	04:52



SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 05/19/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 19030

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:58
INTERNAL STANDARD	VERIFIED	13:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:58
BLANK TEST	0.00	13:59
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:59
BLANK TEST	0.00	14:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:00
BLANK TEST	0.00	14:01







SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 05/11/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 19030

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:28
INTERNAL STANDARD	VERIFIED	12:28
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:29
BLANK TEST	0.00	12:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:30
BLANK TEST	0.00	12:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:31
BLANK TEST	0.00	12:32







SOLUTION CHANGE

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 05/07/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 19030

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:56
INTERNAL STANDARD	VERIFIED	14:56
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:56
BLANK TEST	0.00	14:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:57
BLANK TEST	0.00	14:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:58
BLANK TEST	0.00	14:59







SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 04/28/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 19030

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:10
INTERNAL STANDARD	VERIFIED	01:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:10
BLANK TEST	0.00	01:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:12
BLANK TEST	0.00	01:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:13
BLANK TEST	0.00	01:14





SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 04/20/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 19030

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:50
INTERNAL STANDARD	VERIFIED	21:50
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:50
BLANK TEST	0.00	21:51
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:51
BLANK TEST	0.00	21:52
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:52
BLANK TEST	0.00	21:53







SOLUTION CHANGE

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 04/16/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 19030

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:50
INTERNAL STANDARD	VERIFIED	15:50
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	15:51
BLANK TEST	0.00	15:51
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:52
BLANK TEST	0.00	15:53
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	15:53
BLANK TEST	0.00	15:54





SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 03/24/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 18360

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:59
INTERNAL STANDARD	VERIFIED	07:59
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	07:59
BLANK TEST	0.00	08:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	08:00
BLANK TEST	0.00	08:01
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	08:01
BLANK TEST	0.00	08:02







SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 03/16/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 18360

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:01
INTERNAL STANDARD	VERIFIED	10:01
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:01
BLANK TEST	0.00	10:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:02
BLANK TEST	0.00	10:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:03
BLANK TEST	0.00	10:04







SOLUTION CHANGE

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: [redacted]13306

Date: 03/10/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 18360

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:47
INTERNAL STANDARD	VERIFIED	03:47
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	03:48
BLANK TEST	0.00	03:48
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	03:49
BLANK TEST	0.00	03:50
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	03:50
BLANK TEST	0.00	03:51







SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 03/04/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 18360

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:45
INTERNAL STANDARD	VERIFIED	14:45
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:46
BLANK TEST	0.00	14:47
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	14:47
BLANK TEST	0.00	14:48
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	14:48
BLANK TEST	0.00	14:49







SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 02/23/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 18360

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:50
INTERNAL STANDARD	VERIFIED	23:50
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:50
BLANK TEST	0.00	23:51
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:51
BLANK TEST	0.00	23:52
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:53
BLANK TEST	0.00	23:53





SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 02/16/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 18360

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:39
INTERNAL STANDARD	VERIFIED	21:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:40
BLANK TEST	0.00	21:40
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:41
BLANK TEST	0.00	21:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:42
BLANK TEST	0.00	21:43







TECHNICIAN TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 02/13/2019

Nominal = 0.100 g/210L

BLANK TEST	0.000	14:39
INTERNAL STANDARD	VERIFIED	14:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.098	14:39
BLANK TEST	0.000	14:40
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.098	14:40
BLANK TEST	0.000	14:41
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.098	14:41
BLANK TEST	0.000	14:42
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.098	14:42
BLANK TEST	0.000	14:43

Average = 0.0980
Std Dev = 0.0000



SOLUTION CHANGE

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 02/13/2019

OPERATOR NAME:
THOMAS J ZOVISTOSKI
Permit #: 36771
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 18360

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:02
INTERNAL STANDARD	VERIFIED	14:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:02
BLANK TEST	0.00	14:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:03
BLANK TEST	0.00	14:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:04
BLANK TEST	0.00	14:05







SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 02/09/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 18310

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:27
INTERNAL STANDARD	VERIFIED	22:27
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:27
BLANK TEST	0.00	22:28
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:29
BLANK TEST	0.00	22:29
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:30
BLANK TEST	0.00	22:31







SOLUTION CHANGE

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 02/02/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 18310

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:01
INTERNAL STANDARD	VERIFIED	19:01
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:01
BLANK TEST	0.00	19:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:02
BLANK TEST	0.00	19:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:04
BLANK TEST	0.00	19:04







SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 01/26/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 18230

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:53
INTERNAL STANDARD	VERIFIED	21:53
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	21:54
BLANK TEST	0.00	21:54
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	21:55
BLANK TEST	0.00	21:55
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	21:56
BLANK TEST	0.00	21:57





SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 01/20/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 18230

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:49
INTERNAL STANDARD	VERIFIED	02:49
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:50
BLANK TEST	0.00	02:51
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:51
BLANK TEST	0.00	02:52
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:52
BLANK TEST	0.00	02:53







SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 01/12/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 18230

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:58
INTERNAL STANDARD	VERIFIED	21:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:58
BLANK TEST	0.00	21:59
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:59
BLANK TEST	0.00	22:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:00
BLANK TEST	0.00	22:01







SOLUTION CHANGE

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 01/05/2019

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 18310

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:07
INTERNAL STANDARD	VERIFIED	22:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:08
BLANK TEST	0.00	22:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:09
BLANK TEST	0.00	22:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:10
BLANK TEST	0.00	22:11