SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 12/28/2018

OPERATOR NAME:
KYLE L KOLAR
Permit #: 44063
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 18310

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:08
INTERNAL STANDARD	VERIFIED	17:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:09
BLANK TEST	0.00	17:10
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:10
BLANK TEST	0.00	17:11
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:11
BLANK TEST	0.00	17:13

#691







SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 12/17/2018

OPERATOR NAME:
CORBET SULLIVAN
Permit #: 45898
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: ~~53058~~ 18310

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:45
INTERNAL STANDARD	VERIFIED	16:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:46
BLANK TEST	0.00	16:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	16:47
BLANK TEST	0.00	16:48
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:48
BLANK TEST	0.00	16:49

CS #623







SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 12/10/2018

OPERATOR NAME:
KYLE L KOLAR
Permit #: 44063
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 18310

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:47
INTERNAL STANDARD	VERIFIED	10:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:48
BLANK TEST	0.00	10:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:49
BLANK TEST	0.00	10:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:50
BLANK TEST	0.00	10:51

#691



SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 12/03/2018

OPERATOR NAME:
CHRISTOPHER J MURRAY
Permit #: 46226
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 18310

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:03
INTERNAL STANDARD	VERIFIED	07:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:04
BLANK TEST	0.00	07:05
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:05
BLANK TEST	0.00	07:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:07
BLANK TEST	0.00	07:08







#521

SOLUTION CHANGE

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 12/02/2018

OPERATOR NAME:
KYLE L KOLAR
Permit #: 44063
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 18310

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:16
INTERNAL STANDARD	VERIFIED	14:17
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:17
BLANK TEST	0.00	14:18
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:18
BLANK TEST	0.00	14:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:20
BLANK TEST	0.00	14:21

#691







SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 11/26/2018

OPERATOR NAME:
KYLE L KOLAR
Permit #: 44063
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 18230

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:32
INTERNAL STANDARD	VERIFIED	13:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:33
BLANK TEST	0.00	13:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:34
BLANK TEST	0.00	13:35
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:35
BLANK TEST	0.00	13:36

#691







SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 11/20/2018

OPERATOR NAME:
JENNIFER M NOBLE
Permit #: 44064
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 18230

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:35
INTERNAL STANDARD	VERIFIED	15:36
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:36
BLANK TEST	0.00	15:37
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:37
BLANK TEST	0.00	15:38
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:39
BLANK TEST	0.00	15:40

Jennifer Noble #6087



SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 11/12/2018

OPERATOR NAME:
KYLE L KOLAR
Permit #: 44063
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 18230

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:43
INTERNAL STANDARD	VERIFIED	16:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:43
BLANK TEST	0.00	16:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:44
BLANK TEST	0.00	16:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:46
BLANK TEST	0.00	16:47

KKO #691







SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 11/05/2018

OPERATOR NAME:
JASON R DIAMOND
Permit #: 37496
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 18230

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:44
INTERNAL STANDARD	VERIFIED	17:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:44
BLANK TEST	0.00	17:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:45
BLANK TEST	0.00	17:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:47
BLANK TEST	0.00	17:48

1661







SOLUTION CHANGE

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 11/01/2018

OPERATOR NAME:
CHRISTOPHER J MURRAY
Permit #: 46226
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 18230

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:38
INTERNAL STANDARD	VERIFIED	13:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:38
BLANK TEST	0.00	13:39
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:39
BLANK TEST	0.00	13:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:41
BLANK TEST	0.00	13:42



SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 10/29/2018

OPERATOR NAME:
ETHAN J MACURA
Permit #: 44080
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 18230

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:19
INTERNAL STANDARD	VERIFIED	15:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:19
BLANK TEST	0.00	15:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:21
BLANK TEST	0.00	15:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:22
BLANK TEST	0.00	15:23

#096







SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 10/24/2018

OPERATOR NAME:
CORY L HURLBURT
Permit #: 37017
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 18230

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:57
INTERNAL STANDARD	VERIFIED	08:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:58
BLANK TEST	0.00	08:59
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:59
BLANK TEST	0.00	09:00
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:01
BLANK TEST	0.00	09:02

C. Hurlburt
#257







SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 10/15/2018

OPERATOR NAME:
KATHERINE M PADDOCK
Permit #: 45899
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 18230

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:16
INTERNAL STANDARD	VERIFIED	18:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:16
BLANK TEST	0.00	18:17
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:18
BLANK TEST	0.00	18:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:19
BLANK TEST	0.00	18:20







SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 10/08/2018

OPERATOR NAME:
CORBET W SULLIVAN
Permit #: 45898
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 18230

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:52
INTERNAL STANDARD	VERIFIED	06:52
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:52
BLANK TEST	0.00	06:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:53
BLANK TEST	0.00	06:54
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:55
BLANK TEST	0.00	06:56







SOLUTION CHANGE

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 10/01/2018

OPERATOR NAME:
CHRISTOPHER J MURRAY
Permit #: 46226
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 18230

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:06
INTERNAL STANDARD	VERIFIED	09:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:07
BLANK TEST	0.00	09:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:08
BLANK TEST	0.00	09:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:10
BLANK TEST	0.00	09:11







SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 09/18/2018

OPERATOR NAME:
PAUL M TROTTIER
Permit #: 37497
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 18170

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:55
INTERNAL STANDARD	VERIFIED	08:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:55
BLANK TEST	0.00	08:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:56
BLANK TEST	0.00	08:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:58
BLANK TEST	0.00	08:59







#659

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 09/24/2018

OPERATOR NAME:
CORY L HURLBURT
Permit #: 37017
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 18170

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:10
INTERNAL STANDARD	VERIFIED	07:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:11
BLANK TEST	0.00	07:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:12
BLANK TEST	0.00	07:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:14
BLANK TEST	0.00	07:15

Cit #757







SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 09/10/2018

OPERATOR NAME:
JASON R DIAMOND
Permit #: 37496
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 18170

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:57
INTERNAL STANDARD	VERIFIED	19:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:57
BLANK TEST	0.00	19:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:59
BLANK TEST	0.00	20:00
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:00
BLANK TEST	0.00	20:01

JD 661







SOLUTION CHANGE

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 09/03/2018

OPERATOR NAME:
KATHERINE M PADDOCK
Permit #: 45899
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 18170

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:57
INTERNAL STANDARD	VERIFIED	14:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:57
BLANK TEST	0.00	14:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:59
BLANK TEST	0.00	15:00
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:00
BLANK TEST	0.00	15:01

K Paddock #613







SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 08/27/2018

OPERATOR NAME:
ETHAN J MACURA
Permit #: 44080
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 18170

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:51
INTERNAL STANDARD	VERIFIED	11:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:51
BLANK TEST	0.00	11:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:52
BLANK TEST	0.00	11:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:54
BLANK TEST	0.00	11:55

#684







SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 08/21/2018

OPERATOR NAME:
JOHN K LATOUR
Permit #: 31414
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 18170

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:35
INTERNAL STANDARD	VERIFIED	22:35
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	22:35
BLANK TEST	0.00	22:36
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	22:36
BLANK TEST	0.00	22:37
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	22:38
BLANK TEST	0.00	22:39

#653







SOLUTION CHANGE

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 08/07/2018

OPERATOR NAME:
WILLIAM L HOLCOMB
Permit #: 31942
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 18170

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:02
INTERNAL STANDARD	VERIFIED	14:02
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	14:02

Blank Error



SOLUTION CHANGE

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 08/07/2018

OPERATOR NAME:
WILLIAM L HOLCOMB
Permit #: 31942
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 18170

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:57
INTERNAL STANDARD	VERIFIED	13:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:57

Blank Error



SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 08/14/2018

OPERATOR NAME:
CHRISTOPHER J MURRAY
Permit #: 46226
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 18170

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:03
INTERNAL STANDARD	VERIFIED	07:03
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:04
BLANK TEST	0.00	07:05
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:05
BLANK TEST	0.00	07:06
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:06
BLANK TEST	0.00	07:07

CM #621







SOLUTION CHANGE

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 08/07/2018

OPERATOR NAME:
MATTHEW J LOHRET
Permit #: 28941
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 18170

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:19
INTERNAL STANDARD	VERIFIED	14:19
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:19
BLANK TEST	0.00	14:20
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:21
BLANK TEST	0.00	14:22
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:22
BLANK TEST	0.00	14:23

MJL #619







SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 08/06/2018

OPERATOR NAME:
KATHERINE M PADDOCK
Permit #: 45899
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 18170

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:20
INTERNAL STANDARD	VERIFIED	17:20
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	17:21
BLANK TEST	0.00	17:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:22
BLANK TEST	0.00	17:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:23
BLANK TEST	0.00	17:24

#613
KP







SOLUTION CHANGE

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 08/01/2018

OPERATOR NAME:
CORY L HURLBURT
Permit #: 37017
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 18170

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:44
INTERNAL STANDARD	VERIFIED	07:44
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.10	07:44
BLANK TEST	0.00	07:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:46
BLANK TEST	0.00	07:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:47
BLANK TEST	0.00	07:48

#257







SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 07/30/2018

OPERATOR NAME:
JASON R DIAMOND
Permit #: 37496
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 18100

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:18
INTERNAL STANDARD	VERIFIED	19:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	19:18
BLANK TEST	0.00	19:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:20
BLANK TEST	0.00	19:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:21
BLANK TEST	0.00	19:22







SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 07/23/2018

OPERATOR NAME:
T A EARL
Permit #: 33337
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 18100

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:35
INTERNAL STANDARD	VERIFIED	07:35
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:36
BLANK TEST	0.00	07:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:37
BLANK TEST	0.00	07:38
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:38
BLANK TEST	0.00	07:39







FIRE
#680

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 07/17/2018

OPERATOR NAME:
BRADLEY J HAMILTON
Permit #: 40380
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 18100

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:07
INTERNAL STANDARD	VERIFIED	22:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:07
BLANK TEST	0.00	22:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:09
BLANK TEST	0.00	22:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:10
BLANK TEST	0.00	22:11

BH
#660







SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 07/09/2018

OPERATOR NAME:
ETHAN J MACURA
Permit #: 44080
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 18100

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:27
INTERNAL STANDARD	VERIFIED	23:28
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:28
BLANK TEST	0.00	23:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:29
BLANK TEST	0.00	23:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:31
BLANK TEST	0.00	23:32

E. Macura
#696.







SOLUTION CHANGE

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 07/04/2018

OPERATOR NAME:
CHRISTOPHER J MURRAY
Permit #: 46226
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 18100

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:41
INTERNAL STANDARD	VERIFIED	08:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:41
BLANK TEST	0.00	08:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:43
BLANK TEST	0.00	08:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:44
BLANK TEST	0.00	08:45

C. Murray
#821






0.16
0.14
0.12
0.10
0.08
0.06
0.04
0.02
0.00
32.0
28.0
24.0
20.0
16.0
12.0
8.0
4.0
0.0
0
5
10
15
20
25
30