

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 11/04/2019

OPERATOR NAME:
JASON R DIAMOND
Permit #: 37496
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:52
INTERNAL STANDARD	VERIFIED	14:52
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	14:52
BLANK TEST	0.00	14:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	14:53
BLANK TEST	0.00	14:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:55
BLANK TEST	0.00	14:56

50661

