

SUPERVISOR TEST

WASHINGTON COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106406

Date: 10/21/2019

OPERATOR NAME:
 CHRISTOPHER J MURRAY
 Permit #: 46226
 Operator Agency/Dept: WASHINGTON CO. SO
 Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:27
INTERNAL STANDARD	VERIFIED	11:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:28
BLANK TEST	0.00	11:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:29
BLANK TEST	0.00	11:30
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	11:30
BLANK TEST	0.00	11:31

CM #821

