

SOLUTION CHANGE

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 11/01/2019

OPERATOR NAME:
CORY L HURLBURT
Permit #: 37017
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:44
INTERNAL STANDARD	VERIFIED	07:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:45
BLANK TEST	0.00	07:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:46
BLANK TEST	0.00	07:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:47
BLANK TEST	0.00	07:48

CH # 257

