

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 11/11/2019

OPERATOR NAME:
KYLE L KOLAR
Permit #: 44063
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:20
INTERNAL STANDARD	VERIFIED	13:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:21
BLANK TEST	0.00	13:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:22
BLANK TEST	0.00	13:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	13:24
BLANK TEST	0.00	13:25

R #691

