

SOLUTION CHANGE

 WASHINGTON COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106406

Date: 02/01/2019

OPERATOR NAME:
 JOHN K LATOUR
 Permit #: 31414
 Operator Agency/Dept: WASHINGTON CO. SO
 Reference Standard #: 18360

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:03
INTERNAL STANDARD	VERIFIED	19:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:04
BLANK TEST	0.00	19:05
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:05
BLANK TEST	0.00	19:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:07
BLANK TEST	0.00	19:08

