

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 02/18/2019

OPERATOR NAME:
JASON R DIAMOND
Permit #: 37496
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 18360

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:14
INTERNAL STANDARD	VERIFIED	14:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:14
BLANK TEST	0.00	14:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:16
BLANK TEST	0.00	14:17
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:17
BLANK TEST	0.00	14:18

