

SUPERVISOR TEST

WASHINGTON COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106406

Date: 03/18/2019

OPERATOR NAME:
 KATHERINE M PADDOCK
 Permit #: 45899
 Operator Agency/Dept: WASHINGTON CO. SO
 Reference Standard #: 18360

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:50
INTERNAL STANDARD	VERIFIED	08:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:50
BLANK TEST	0.00	08:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:52
BLANK TEST	0.00	08:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:53
BLANK TEST	0.00	08:54

Handwritten: KP #113

