SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 03/04/2019

OPERATOR NAME:
CORBET W SULLIVAN
Permit #: 45898
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 18360

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:51
INTERNAL STANDARD	VERIFIED	06:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:52
BLANK TEST	0.00	06:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:53
BLANK TEST	0.00	06:54
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:54
BLANK TEST	0.00	06:55

CS
#623