

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 05/10/2019

OPERATOR NAME:
KYLE L KOLAR
Permit #: 44063
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 19030

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:10
INTERNAL STANDARD	VERIFIED	14:10
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:11
BLANK TEST	0.00	14:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:12
BLANK TEST	0.00	14:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:13
BLANK TEST	0.00	14:14

[Handwritten signature] #691

