

SOLUTION CHANGE

WASHINGTON COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106406

Date: 06/03/2019

OPERATOR NAME:  
WILLIAM L HOLCOMB  
Permit #: 31942  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:44
INTERNAL STANDARD	VERIFIED	15:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:44
BLANK TEST	0.00	15:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:46
BLANK TEST	0.00	15:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:47
BLANK TEST	0.00	15:48

Sgt. W. Holcomb #679

