

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 07/22/2019

OPERATOR NAME:
KYLE L KOLAR
Permit #: 44063
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:44
INTERNAL STANDARD	VERIFIED	22:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:45
BLANK TEST	0.00	22:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:46
BLANK TEST	0.00	22:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:48
BLANK TEST	0.00	22:49

Q #691

