

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 07/08/2019

OPERATOR NAME:
KATHERINE M PADDOCK
Permit #: 45899
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:18
INTERNAL STANDARD	VERIFIED	00:18
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:18
BLANK TEST	0.00	00:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:20
BLANK TEST	0.00	00:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:21
BLANK TEST	0.00	00:22

KMP #1013

