

SUPERVISOR TEST

WASHINGTON COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106406

Date: 08/19/2019

OPERATOR NAME:  
KYLE L KOLAR  
Permit #: 44063  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:55
INTERNAL STANDARD	VERIFIED	16:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	16:55
BLANK TEST	0.00	16:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	16:57
BLANK TEST	0.00	16:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	16:58
BLANK TEST	0.00	16:59

*W #691*

