

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 08/05/2019

OPERATOR NAME:
CORY L HURLBURT
Permit #: 37017
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 19130

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 08:55 |
| INTERNAL STANDARD | VERIFIED | 08:55 |
| REFERENCE STANDARD TEMP | 34.01c | |
| REFERENCE STANDARD | 0.10 | 08:56 |
| BLANK TEST | 0.00 | 08:57 |
| REFERENCE STANDARD TEMP | 34.01c | |
| REFERENCE STANDARD | 0.10 | 08:57 |
| BLANK TEST | 0.00 | 08:58 |
| REFERENCE STANDARD TEMP | 34.01c | |
| REFERENCE STANDARD | 0.10 | 08:58 |
| BLANK TEST | 0.00 | 09:00 |

Cory Hurlburt # 257

