

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 09/16/2019

OPERATOR NAME:
KYLE L KOLAR
Permit #: 44063
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:03
INTERNAL STANDARD	VERIFIED	08:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:03
BLANK TEST	0.00	08:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:05
BLANK TEST	0.00	08:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:06
BLANK TEST	0.00	08:07

W #691

