

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 09/23/2019

OPERATOR NAME:
WILLIAM L HOLCOMB
Permit #: 31942
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:54
INTERNAL STANDARD	VERIFIED	17:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:55
BLANK TEST	0.00	17:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:56
BLANK TEST	0.00	17:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:58
BLANK TEST	0.00	17:59

Sgt. W. Holcomb #679

