

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 01/21/2019

OPERATOR NAME:
MATTHEW K JACKSON
Permit #: 40078
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 18310

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:25
INTERNAL STANDARD	VERIFIED	18:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:26
BLANK TEST	0.00	18:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:27
BLANK TEST	0.00	18:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:28
BLANK TEST	0.00	18:29

YAP
1693

