

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 01/07/2019

OPERATOR NAME:
D TODD RICHE
Permit #: 33411
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 18310

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:31
INTERNAL STANDARD	VERIFIED	11:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:32
BLANK TEST	0.00	11:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:33
BLANK TEST	0.00	11:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:34
BLANK TEST	0.00	11:35

D. Riche

