

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 11/11/2019

OPERATOR NAME:
MICHAEL WEBER
Permit #: 46067
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:51
INTERNAL STANDARD	VERIFIED	23:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:51
BLANK TEST	0.00	23:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:52
BLANK TEST	0.00	23:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:54
BLANK TEST	0.00	23:54

