

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 04/15/2019

OPERATOR NAME:
MATTHEW K JACKSON
Permit #: 40078
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 19030

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:01
INTERNAL STANDARD	VERIFIED	14:01
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:01
BLANK TEST	0.00	14:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:03
BLANK TEST	0.00	14:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:04
BLANK TEST	0.00	14:05

MAT # 693

