

SUPERVISOR TEST

WASHINGTON CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100284




Date: 05/13/2019

OPERATOR NAME:  
CORI J WINCH  
Permit #: 44971  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 1138

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:23
INTERNAL STANDARD	VERIFIED	06:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:24
BLANK TEST	0.00	06:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:25
BLANK TEST	0.00	06:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:26
BLANK TEST	0.00	06:27

#697



SUPERVISOR TEST

WASHINGTON CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100284

Date: 05/13/2019

OPERATOR NAME:  
D TODD RICHE  
Permit #: 33411  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 19030

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:17
INTERNAL STANDARD	VERIFIED	07:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:17
BLANK TEST	0.00	07:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:18
BLANK TEST	0.00	07:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:19
BLANK TEST	0.00	07:20

610

