

SUPERVISOR TEST

-----  
WASHINGTON CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100284  
-----

Date: 05/20/2019

OPERATOR NAME:  
ROBERT J SULLIVAN  
Permit #: 34106  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 19030

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:34
INTERNAL STANDARD	VERIFIED	03:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	03:34
BLANK TEST	0.00	03:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	03:35
BLANK TEST	0.00	03:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	03:36
BLANK TEST	0.00	03:37

