

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 05/27/2019

OPERATOR NAME:
D TODD RICHE
Permit #: 33411
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 19030

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 15:56 |
| INTERNAL STANDARD | VERIFIED | 15:56 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 15:56 |
| BLANK TEST | 0.00 | 15:57 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 15:58 |
| BLANK TEST | 0.00 | 15:58 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 15:59 |
| BLANK TEST | 0.00 | 15:59 |

