

SUPERVISOR TEST

WASHINGTON CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100284

Date: 06/25/2019

OPERATOR NAME:  
MATTHEW K JACKSON  
Permit #: 40078  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:40
INTERNAL STANDARD	VERIFIED	16:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:40
BLANK TEST	0.00	16:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:41
BLANK TEST	0.00	16:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:42
BLANK TEST	0.00	16:43

693472

