

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 08/05/2019

OPERATOR NAME:
D TODD RICHE
Permit #: 33411
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:35
INTERNAL STANDARD	VERIFIED	09:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:36
BLANK TEST	0.00	09:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:37
BLANK TEST	0.00	09:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:38
BLANK TEST	0.00	09:39

Handwritten signature

