

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 09/09/2019

OPERATOR NAME:
MATTHEW K JACKSON
Permit #: 40078
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:51
INTERNAL STANDARD	VERIFIED	17:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:51
BLANK TEST	0.00	17:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:52
BLANK TEST	0.00	17:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:54
BLANK TEST	0.00	17:54


