



**State
Police**

**Forensic
Investigation
Center**

ANDREW M. CUOMO

Governor

KEITH M. CORLETT

Acting Superintendent

RAY A. WICKENHEISER

Director

0.10% SIMULATOR SOLUTION RECORD

Report Date: February 7, 2019
Manufacturer: Guth Laboratories, Inc.
Lot Number: 19030
Expiration Date: 7/21/2019
Analysis Results: 120 mg Ethyl Alcohol / 100 mL of solution

I, Carol S Miller, tested simulator solution lot number 19030 by headspace gas chromatography and have determined that it contains the appropriate concentration of ethyl alcohol. This solution is hereby approved for use.

When this simulator solution is used with a properly operating breath testing instrument, it will provide a value of 0.10% within acceptable limits.

I, Carol S Miller, made the entries contained in this record at the time of each event recorded, or within a reasonable time thereafter.

Performed by:

Digitally signed under
ESRA by Carol S Miller on
2019.02.07 14:22:48 -05'00'

Carol S Miller
Forensic Scientist III
Breath Analysis Technical Supervisor/Operator #43419

Reviewed by:

Digitally signed under
ESRA by Carrie A Kirkton
on
2019.02.12 08:05:13 -05'00'

Carrie A Kirkton
Supervisor of Forensic Services
Breath Analysis Technical Supervisor