

SUPERVISOR TEST

-----  
 HUDSON FALLS  
 VILLAGE POLICE DEPT.  
 DATAMASTER dmt: 102408  
 -----

Date: 11/28/2019

OPERATOR NAME:  
 RICHARD J ARNOLD  
 Permit #: 33860  
 Operator Agency/Dept: HUDSON FALLS VIL. PD  
 Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:56
INTERNAL STANDARD	VERIFIED	21:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:56
BLANK TEST	0.00	21:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:57
BLANK TEST	0.00	21:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:59
BLANK TEST	0.00	21:59

