

SUPERVISOR TEST

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

ce: 07/17/2019

OPERATOR NAME:
RICHARD J ARNOLD
Permit #: 33860
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 19090

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 02:29 |
| INTERNAL STANDARD | VERIFIED | 02:29 |
| REFERENCE STANDARD TEMP | 33.98c | |
| REFERENCE STANDARD | 0.10 | 02:30 |
| BLANK TEST | 0.00 | 02:31 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 02:31 |
| BLANK TEST | 0.00 | 02:32 |
| REFERENCE STANDARD TEMP | 34.01c | |
| REFERENCE STANDARD | 0.10 | 02:32 |
| BLANK TEST | 0.00 | 02:33 |

