

SUPERVISOR TEST

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 08/21/2019

OPERATOR NAME:
BRAD M LYON
Permit #: 44978
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 19130

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 01:33 |
| INTERNAL STANDARD | VERIFIED | 01:33 |
| REFERENCE STANDARD TEMP | 33.99c | |
| REFERENCE STANDARD | 0.10 | 01:34 |
| BLANK TEST | 0.00 | 01:35 |
| REFERENCE STANDARD TEMP | 33.99c | |
| REFERENCE STANDARD | 0.10 | 01:35 |
| BLANK TEST | 0.00 | 01:36 |
| REFERENCE STANDARD TEMP | 33.97c | |
| REFERENCE STANDARD | 0.10 | 01:36 |
| BLANK TEST | 0.00 | 01:37 |

