

SUPERVISOR TEST

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 10/16/2019

OPERATOR NAME:
RICHARD J ARNOLD
Permit #: 33860
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:36
INTERNAL STANDARD	VERIFIED	19:36
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	19:36
BLANK TEST	0.00	19:37
REFERENCE STANDARD TEMP	34.05c	
REFERENCE STANDARD	0.09	19:38
BLANK TEST	0.00	19:38
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:39
BLANK TEST	0.00	19:40

