

SOLUTION CHANGE

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 10/30/2019

OPERATOR NAME:
RICHARD J ARNOLD
Permit #: 33860
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:13
INTERNAL STANDARD	VERIFIED	21:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:13
BLANK TEST	0.00	21:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:14
BLANK TEST	0.00	21:15
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.10	21:16
BLANK TEST	0.00	21:16

