

SUPERVISOR TEST




-----  
FORT EDWARD  
VILLAGE POLICE DEPT.  
DATAMASTER dmt: 115708  
-----

Date: 01/13/2020

OPERATOR NAME:  
SGT DEAN E WATKINS  
Permit #: 39523  
Operator Agency/Dept: FORT EDWARD VILL. PD  
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:13
INTERNAL STANDARD	VERIFIED	09:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:13
BLANK TEST	0.00	09:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:15
BLANK TEST	0.00	09:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:16
BLANK TEST	0.00	09:17



SOLUTION CHANGE




-----  
FORT EDWARD  
VILLAGE POLICE DEPT.  
DATAMASTER dmt: 115708  
-----

Date: 01/06/2020

OPERATOR NAME:  
SGT DEAN E WATKINS  
Permit #: 39523  
Operator Agency/Dept: FORT EDWARD VILL. PD  
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:31
INTERNAL STANDARD	VERIFIED	20:31
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:31
BLANK TEST	0.00	20:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:33
BLANK TEST	0.00	20:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:34
BLANK TEST	0.00	20:35



SUPERVISOR TEST




-----  
FORT EDWARD  
VILLAGE POLICE DEPT.  
DATAMASTER dmt: 115708  
-----

Date: 12/30/2019

OPERATOR NAME:  
SGT DEAN E WATKINS  
Permit #: 39523  
Operator Agency/Dept: FORT EDWARD VILL. PD  
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:10
INTERNAL STANDARD	VERIFIED	20:10
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:11
BLANK TEST	0.00	20:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:12
BLANK TEST	0.00	20:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:13
BLANK TEST	0.00	20:14



SUPERVISOR TEST




-----  
FORT EDWARD  
VILLAGE POLICE DEPT.  
DATAMASTER dmt: 115708  
-----

Date: 12/23/2019

OPERATOR NAME:  
SGT DEAN E WATKINS  
Permit #: 39523  
Operator Agency/Dept: FORT EDWARD VILL. PD  
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:49
INTERNAL STANDARD	VERIFIED	08:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:49
BLANK TEST	0.00	08:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:50
BLANK TEST	0.00	08:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:52
BLANK TEST	0.00	08:53



SUPERVISOR TEST

-----  
FORT EDWARD  
VILLAGE POLICE DEPT.  
DATAMASTER dmt: 115708  
-----

Date: 12/17/2019

OPERATOR NAME:  
SGT DEAN E WATKINS  
Permit #: 39523  
Operator Agency/Dept: FORT EDWARD VILL. PD  
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:21
INTERNAL STANDARD	VERIFIED	10:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:21
BLANK TEST	0.00	10:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:23
BLANK TEST	0.00	10:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:24
BLANK TEST	0.00	10:25

