

SUPERVISOR TEST

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 01/14/2020

OPERATOR NAME:
RICHARD J ARNOLD
Permit #: 33860
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:41
INTERNAL STANDARD	VERIFIED	23:41
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:41
BLANK TEST	0.00	23:42
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:43
BLANK TEST	0.00	23:43
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:44
BLANK TEST	0.00	23:45

