

SUPERVISOR TEST

-----  
HUDSON FALLS  
VILLAGE POLICE DEPT.  
DATAMASTER dmt: 102408  
-----

Date: 01/02/2020

OPERATOR NAME:  
BRAD M LYON  
Permit #: 44978  
Operator Agency/Dept: HUDSON FALLS VIL. PD  
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:24
INTERNAL STANDARD	VERIFIED	02:24
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	02:24
BLANK TEST	0.00	02:25
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	02:25
BLANK TEST	0.00	02:26
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	02:27
BLANK TEST	0.00	02:28

