

SOLUTION CHANGE

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 01/08/2020

OPERATOR NAME:
RICHARD J ARNOLD
Permit #: 33860
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:55
INTERNAL STANDARD	VERIFIED	00:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:56
BLANK TEST	0.00	00:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:57
BLANK TEST	0.00	00:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:58
BLANK TEST	0.00	00:59

