

SOLUTION CHANGE

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 01/01/2020

OPERATOR NAME:
CHRISTOPHER J MURRAY
Permit #: 46226
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:03
INTERNAL STANDARD	VERIFIED	09:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:03
BLANK TEST	0.00	09:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:05
BLANK TEST	0.00	09:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:06
BLANK TEST	0.00	09:07

CM
#821

