

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 01/06/2020

OPERATOR NAME:
KYLE L KOLAR
Permit #: 44063
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:17
INTERNAL STANDARD	VERIFIED	19:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	19:17
BLANK TEST	0.00	19:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	19:19
BLANK TEST	0.00	19:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	19:20
BLANK TEST	0.00	19:21

Q#691

lll

