

SUPERVISOR TEST

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 01/22/2020

OPERATOR NAME:
RICHARD J ARNOLD
Permit #: 33860
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:02
INTERNAL STANDARD	VERIFIED	19:02
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:02
BLANK TEST	0.00	19:03
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:03
BLANK TEST	0.00	19:04
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:04
BLANK TEST	0.00	19:05

