SUPERVISOR TEST

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 01/15/2020

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:02
INTERNAL STANDARD	VERIFIED	07:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:02
BLANK TEST	0.00	07:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:03
BLANK TEST	0.00	07:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:04
BLANK TEST dmt: 126706	0.00	07:05

Date: 01/15/2020

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:02
INTERNAL STANDARD	VERIFIED	07:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:02
BLANK TEST	0.00	07:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:03
BLANK TEST	0.00	07:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:04
BLANK TEST	0.00	07:05







SUPERVISOR TEST

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 01/06/2020

OPERATOR NAME:
PAUL-HENRY N LAPORTE
Permit #: 47733
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:56
INTERNAL STANDARD	VERIFIED	13:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:57
BLANK TEST	0.00	13:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:58
BLANK TEST	0.00	13:59
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:59
BLANK TEST	0.00	14:00



SOLUTION CHANGE

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 01/03/2020

OPERATOR NAME:
CHRISTOPHER R DAVIS-FLYNN
Permit #: 41324
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:21
INTERNAL STANDARD	VERIFIED	07:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:21
BLANK TEST	0.00	07:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:22
BLANK TEST	0.00	07:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:23
BLANK TEST	0.00	07:24



SUPERVISOR TEST

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 12/31/2019

OPERATOR NAME:
CHRISTOPHER R DAVIS-FLYNN
Permit #: 41324
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:25
INTERNAL STANDARD	VERIFIED	15:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:25
BLANK TEST	0.00	15:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:26
BLANK TEST	0.00	15:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:27
BLANK TEST	0.00	15:28



SUPERVISOR TEST

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 12/23/2019

OPERATOR NAME:
RYAN M BUELL
Permit #: 46581
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:08
INTERNAL STANDARD	VERIFIED	07:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:08
BLANK TEST	0.00	07:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:09
BLANK TEST	0.00	07:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:10
BLANK TEST	0.00	07:11



SUPERVISOR TEST

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 12/16/2019

OPERATOR NAME:
CHRISTOPHER R DAVIS-FLYNN
Permit #: 41324
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:24
INTERNAL STANDARD	VERIFIED	10:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:25
BLANK TEST	0.00	10:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:26
BLANK TEST	0.00	10:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:27
BLANK TEST	0.00	10:28







SUPERVISOR TEST

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 12/09/2019

OPERATOR NAME:
RYAN M BUELL
Permit #: 46581
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:54
INTERNAL STANDARD	VERIFIED	12:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:54
BLANK TEST	0.00	12:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:55
BLANK TEST	0.00	12:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:56
BLANK TEST	0.00	12:57







SOLUTION CHANGE

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 12/03/2019

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:12
INTERNAL STANDARD	VERIFIED	08:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:12
BLANK TEST	0.00	08:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:13
BLANK TEST	0.00	08:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:14
BLANK TEST	0.00	08:15







SUPERVISOR TEST

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 11/25/2019

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:02
INTERNAL STANDARD	VERIFIED	07:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:03
BLANK TEST	0.00	07:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:04
BLANK TEST	0.00	07:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:05
BLANK TEST	0.00	07:06







SUPERVISOR TEST

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 11/18/2019

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:07
INTERNAL STANDARD	VERIFIED	07:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:08
BLANK TEST	0.00	07:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:09
BLANK TEST	0.00	07:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:10
BLANK TEST	0.00	07:11



SUPERVISOR TEST

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 11/11/2019

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:30
INTERNAL STANDARD	VERIFIED	18:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:30
BLANK TEST	0.00	18:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:31
BLANK TEST	0.00	18:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:32
BLANK TEST dmt: 126706	0.00	18:33







SUPERVISOR TEST

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 11/04/2019

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:16
INTERNAL STANDARD	VERIFIED	07:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:17
BLANK TEST	0.00	07:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:18
BLANK TEST	0.00	07:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:19
BLANK TEST	0.00	07:20







SOLUTION CHANGE

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 11/01/2019

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:07
INTERNAL STANDARD	VERIFIED	13:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:08
BLANK TEST	0.00	13:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:09
BLANK TEST	0.00	13:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:10
BLANK TEST dmt: 126706	0.00	13:11







SUPERVISOR TEST

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 10/28/2019

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:37
INTERNAL STANDARD	VERIFIED	16:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	16:37
BLANK TEST	0.00	16:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:38
BLANK TEST	0.00	16:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	16:39
BLANK TEST	0.00	16:40



SUPERVISOR TEST

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 10/22/2019

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:08
INTERNAL STANDARD	VERIFIED	18:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	18:09
BLANK TEST	0.00	18:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	18:10
BLANK TEST	0.00	18:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:11
BLANK TEST	0.00	18:12



SUPERVISOR TEST

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 10/15/2019

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:12
INTERNAL STANDARD	VERIFIED	06:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:13
BLANK TEST	0.00	06:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:14
BLANK TEST	0.00	06:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:15
BLANK TEST	0.00	06:16







SUPERVISOR TEST

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 10/07/2019

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:07
INTERNAL STANDARD	VERIFIED	07:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:07
BLANK TEST	0.00	07:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:08
BLANK TEST	0.00	07:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:09
BLANK TEST	0.00	07:10



SOLUTION CHANGE

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 10/01/2019

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:53
INTERNAL STANDARD	VERIFIED	09:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:53
BLANK TEST	0.00	09:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:54
BLANK TEST	0.00	09:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:55
BLANK TEST	0.00	09:56







SUPERVISOR TEST

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 09/30/2019

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:14
INTERNAL STANDARD	VERIFIED	07:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:14
BLANK TEST	0.00	07:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:16
BLANK TEST	0.00	07:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:17
BLANK TEST	0.00	07:17







SUPERVISOR TEST

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 09/23/2019

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:35
INTERNAL STANDARD	VERIFIED	08:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:35
BLANK TEST	0.00	08:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:36
BLANK TEST	0.00	08:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:38
BLANK TEST	0.00	08:38







SUPERVISOR TEST

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 09/16/2019

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:12
INTERNAL STANDARD	VERIFIED	07:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:12
BLANK TEST	0.00	07:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:14
BLANK TEST	0.00	07:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:15
BLANK TEST	0.00	07:15



SUPERVISOR TEST

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 09/09/2019

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:32
INTERNAL STANDARD	VERIFIED	07:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:32
BLANK TEST	0.00	07:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:33
BLANK TEST	0.00	07:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:34
BLANK TEST dmt: 126706	0.00	07:35







SOLUTION CHANGE

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 09/03/2019

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:39
INTERNAL STANDARD	VERIFIED	10:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:39
BLANK TEST	0.00	10:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:40
BLANK TEST	0.00	10:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:41
BLANK TEST dmt: 126706	0.00	10:42

Date: 09/03/2019

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST 0.00 10:39
INTERNAL STANDARD VERIFIED 10:39
REFERENCE STANDARD TEMP 34.00c
REFERENCE STANDARD 0.10 10:39
BLANK TEST 0.00 10:40
REFERENCE STANDARD TEMP 34.00c
REFERENCE STANDARD 0.10 10:40
BLANK TEST 0.00 10:41
REFERENCE STANDARD TEMP 34.00c
REFERENCE STANDARD 0.10 10:41
BLANK TEST 0.00 10:42







SUPERVISOR TEST

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 08/26/2019

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:19
INTERNAL STANDARD	VERIFIED	07:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:20
BLANK TEST	0.00	07:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:21
BLANK TEST	0.00	07:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:22
BLANK TEST dmt: 126706	0.00	07:23







SUPERVISOR TEST

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 08/19/2019

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:10
INTERNAL STANDARD	VERIFIED	07:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:11
BLANK TEST	0.00	07:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:12
BLANK TEST	0.00	07:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:13
BLANK TEST	0.00	07:14







SUPERVISOR TEST

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 08/12/2019

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:22
INTERNAL STANDARD	VERIFIED	08:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:23
BLANK TEST	0.00	08:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:24
BLANK TEST	0.00	08:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:25
BLANK TEST dmt: 126706	0.00	08:26



SUPERVISOR TEST

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 08/06/2019

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:17
INTERNAL STANDARD	VERIFIED	10:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:17
BLANK TEST	0.00	10:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:18
BLANK TEST	0.00	10:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:19
BLANK TEST	0.00	10:20







SOLUTION CHANGE

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 08/01/2019

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:06
INTERNAL STANDARD	VERIFIED	10:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:07
BLANK TEST	0.00	10:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:08
BLANK TEST	0.00	10:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:09
BLANK TEST	0.00	10:10



Date: 08/01/2019

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:06
INTERNAL STANDARD	VERIFIED	10:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:07
BLANK TEST	0.00	10:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:08
BLANK TEST	0.00	10:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:09
BLANK TEST	0.00	10:10





SUPERVISOR TEST

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 07/30/2019

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:25
INTERNAL STANDARD	VERIFIED	19:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:25
BLANK TEST	0.00	19:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:26
BLANK TEST	0.00	19:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:28
BLANK TEST	0.00	19:28



Date: 07/30/2019

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:25
INTERNAL STANDARD	VERIFIED	19:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:25
BLANK TEST	0.00	19:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:26
BLANK TEST	0.00	19:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:28
BLANK TEST	0.00	19:28





SUPERVISOR TEST

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 07/22/2019

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:31
INTERNAL STANDARD	VERIFIED	07:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:31
BLANK TEST	0.00	07:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:32
BLANK TEST	0.00	07:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:33
BLANK TEST	0.00	07:34







SUPERVISOR TEST

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 07/15/2019

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:26
INTERNAL STANDARD	VERIFIED	09:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:26
BLANK TEST	0.00	09:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:27
BLANK TEST	0.00	09:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:29
BLANK TEST	0.00	09:29







SUPERVISOR TEST

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 07/08/2019

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:18
INTERNAL STANDARD	VERIFIED	17:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:18
BLANK TEST	0.00	17:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:19
BLANK TEST	0.00	17:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:20
BLANK TEST	0.00	17:21







SOLUTION CHANGE

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 07/01/2019

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:05
INTERNAL STANDARD	VERIFIED	09:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:05
BLANK TEST	0.00	09:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:06
BLANK TEST	0.00	09:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:07
BLANK TEST	0.00	09:08