

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 01/27/2020

OPERATOR NAME:
D TODD RICHE
Permit #: 33411
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:43
INTERNAL STANDARD	VERIFIED	06:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:44
BLANK TEST	0.00	06:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:45
BLANK TEST	0.00	06:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:46
BLANK TEST	0.00	06:47

DTR
6/10

