

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 01/25/2020

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:52
INTERNAL STANDARD	VERIFIED	21:52
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	21:53
BLANK TEST	0.00	21:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:54
BLANK TEST	0.00	21:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:55
BLANK TEST	0.00	21:56

