

SOLUTION CHANGE

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 02/02/2020

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:01
INTERNAL STANDARD	VERIFIED	03:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:01
BLANK TEST	0.00	03:02
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:02
BLANK TEST	0.00	03:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:03
BLANK TEST	0.00	03:04

