

SOLUTION CHANGE

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 02/02/2020

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:45
INTERNAL STANDARD	VERIFIED	20:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	20:45
BLANK TEST	0.00	20:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	20:46
BLANK TEST	0.00	20:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	20:47
BLANK TEST dmt: 126706	0.00	20:48

Date: 02/02/2020

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:45
INTERNAL STANDARD	VERIFIED	20:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	20:45
BLANK TEST	0.00	20:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	20:46
BLANK TEST	0.00	20:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	20:47
BLANK TEST	0.00	20:48

