

SUPERVISOR TEST

WASHINGTON COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106406

Date: 02/03/2020

OPERATOR NAME:
 CHRISTOPHER J MURRAY
 Permit #: 46226
 Operator Agency/Dept: WASHINGTON CO. SO
 Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:18
INTERNAL STANDARD	VERIFIED	08:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:19
BLANK TEST	0.00	08:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:20
BLANK TEST	0.00	08:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	08:21
BLANK TEST	0.00	08:22

CM #821

