

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 02/08/2020

OPERATOR NAME:
BRYAN J GRECO
Permit #: 44077
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:07
INTERNAL STANDARD	VERIFIED	06:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:08
BLANK TEST	0.00	06:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:09
BLANK TEST	0.00	06:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:10
BLANK TEST	0.00	06:11

