

SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 02/10/2020

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:00
INTERNAL STANDARD	VERIFIED	20:00
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:01
BLANK TEST	0.00	20:02
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:02
BLANK TEST	0.00	20:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:03
BLANK TEST	0.00	20:04

